

Funke Medical GmbH – Declaration - Bacterial testresult

MANUFACTURER'S CERTIFICATE REGARDING ASTM F1670 & ASTM F1671 Resistance to penetration by synthetic blood / by blood-borne pathogens on the mattress covers from Funke Medical GmbH.

Manufacturer of the blue mattress covers for Funke Medical:

CENTEXBEL-VERVIERS
Avenue du Parc 38
BE-4650 Herve (Chaineux)

Date: 09 March 2026

With this document we state that our supplier of the cover material of the Funke Medical mattresses have provided proof of the test result on:

- **ASTM F1670 (2008) Resistance to penetration by synthetic blood**
- **ASTM F1671 (2013) Resistance to penetration by blood-borne pathogens – Test method using Phi-X174 bacteriophage**

We have provided a separate copy of the full test report from our supplier.

The result of the test is:

Test specimen	Pass / Fail results
#1	PASS
#2	PASS
#3	PASS

Pass = no penetration of synthetic blood

Fail = synthetic blood penetration

Name: Rick van Eijken
Title: Director Sales International
Company Name: Funke Medical GmbH

Contact details: Rick van Eijken
Email: r.vaneijken@funke-medical.de
Telephone: +49 2865 266 93 50

Signature:

